

July 11, 2022

RE: Shelf-Life Expiration Update of the GenBody COVID-19 Ag Kit

To Whom It May Concern:

This letter is to inform you that after review of ongoing real-time stability testing data performed by GenBody, Inc., FDA granted an extension of the shelf-life expiration date for the GenBody COVID-19 Ag test from 12 months to 21 months when stored at 2°- 30°C. As a result, the GenBody COVID-19 Ag tests with printed expiration dates between October 2022 and April 2023 now have later expiration dates than those that are currently printed.

In the table below, you will find the Lot numbers affected, their original (printed) expiration dates, and the updated expiration date at the right.

Lot No	Printed Expiration Date	Updated Expiration Date
FGFB04221	04/03/23	01/03/24
FGFN28221	03/27/23	12/27/23
FGFN14221	03/14/23	12/14/23
FVFE16221	02/15/23	11/15/23
FVFM15211	11/14/22	8/14/23
FVFM08211	11/07/22	8/07/23
FVFM01211	10/31/22	7/31/23
FVFA25211	10/24/22	7/24/23
FVFA18211	10/17/22	7/17/23
FVFA10211	10/09/22	7/09/23

For all of the lots designated in the table above, this letter should be provided to all end users with previously distributed product, as well as upon future distribution of the designated lots. Per GenBody's communication with FDA, all test devices in the lots designated above may be used until the updated expiration date without any additional labeling required.

Sincerely,



Jin Soo Kim
Chief Executive Officer
GenBody Inc.

cc: Chomkyu Chong, CTO, GenBody, Inc.
David Yoo, CEO, Kwell Laboratories, LLC (US Agent)